

Reference Material Producer Name : Quality Analysis Department,
PTT Oil and Retail Business Public Company Limited

Address : 555 Ardnarong Road, Klongtoey, Bangkok 10260

Accreditation Number : RMP – 0001

Contact Person : Mr. Pongsak Phttarametevinyu
Phone : 0-2537-2000 to 5457, 0-2537-2000 to 5145
Fax : 0-2537-2000 to 5334, 0-2537-2000 to 7149
E-mail : Pongsak.p@pttplc.com

Initial Issue Date : 3rd April 2014

Issue Date : 2nd May 2022

Expired date : 1st May 2026

Standard : ISO/IEC 17034:2016

Reference No. : 0303/6592

Scope of Reference Material Producer Accreditation

Reference Material Producer Name : Quality Analysis Department,
 PTT Oil and Retail Business Public Company Limited

Address : 555 Ardnarong Road, Klongtoey, Bangkok 10260

Accreditation Number : RMP - 0001

Item Number	Category of reference material	Range of property value	Expanded uncertainty (Confidence level approximately 95%)	Test method/ Technique used
1	Petroleum products - Flash point of Diesel fuel	52 °C to 75 °C	± 1.2 °C	ASTM D93 Procedure A (Pensky-Martens closed cup tester)

Initial Issue Date 3rd April 2014

Issue Number 6

Bureau of Laboratory Accreditation, Department of Science Service, Ministry of Higher Education, Science, Research and Innovation