Name of Laboratory : Laboratory of Siam Winery Co., Ltd.

: 9/2 Moo 3, Tambon Bangtorad, Amphoe Mueang, Address

Changwat Samut Sakhon 74000

Accreditation Number: Testing - 0031 : Miss. Kanokwan Hosakul Contact Person

Phone: 0-3484-5334-43 Ext. 604

Fax: 0-3484-5334-43 Ext. 804

E-mail: kanokwan.h@Siamwinery.com

Initial Issue Date : 26th May 2009

: 15th March 2023 Issue Date

: 14th March 2027 Expired date Standard : ISO/IEC 17025:2017

Reference No.

: 0303/4174

Reference No.: 0303/4174

Scope of Testing Laboratory Accreditation

Laboratory Name : Laboratory of Siam Winery Co., Ltd.

Address : 9/2 Moo 3, Tambon Bangtorad, Amphoe Mueang,

Changwat Samut Sakhon 74000

Accreditation Number : Testing - 0031

Laboratory Status : $oxed{\square}$ Permanent $oxed{\square}$ Site $oxed{\square}$ Temporary $oxed{\square}$ Mobile

Item	Test Material /	Test Item /	Test Method /
Number	Product	Range of Testing	Technique Used
1	Wine and wine	- Alcohol	In - house method : S-QUC-WT01
	based products	2.5 % v/v to 15.0 % v/v	based on Analytica EBC, 2008,
			part 9.2.6
		- Density	In - house method : S-QUC-WT01
		at 20 ℃	based on ASTM : D4052-22
		0.5000 g/cm ³ to 1.5000 g/cm ³	
		- Specific gravity	
		at 20/20 ℃	
		0.5000 to 1.5000	
		- Total acidity	In - house method : S-QUC-WT02
		(Calculated as tartaric acid)	based on AOAC Official Methods
		0.5 g/L to 15.0 g/L	of Analysis, 21 st ed., 2019,
			method 962.12

Initial Issue Date 26th May 2009

Issue Number 7

Bureau of Laboratory Accreditation, Department of Science Service, Ministry of Higher Education, Science, Research and Innovation

LA-F-31-9/11-19 page 1/2

Reference No.: 0303/4174

Scope of Testing Laboratory Accreditation

Laboratory Name : Laboratory of Siam Winery Co., Ltd.

Address : 9/2 Moo 3, Tambon Bangtorad, Amphoe Mueang,

Changwat Samut Sakhon 74000

Accreditation Number : Testing - 0031

Laboratory Status : $oxed{\square}$ Permanent $oxed{\square}$ Site $oxed{\square}$ Temporary $oxed{\square}$ Mobile

Item	Test Material /	Test Item /	Test Method /	
Number	Product	Range of Testing	Technique Used	
1	Wine and wine	- pH	In - house method : S-QUC-WT02	
(cont.)	based products	2.0 to 5.0	based on AOAC Official Methods	
			of Analysis, 21 st ed., 2019,	
			method 960.19	
1			I	

Initial Issue Date 26th May 2009

Issue Number 7

Bureau of Laboratory Accreditation, Department of Science Service, Ministry of Higher Education, Science, Research and Innovation

LA-F-31-9/11-19 page 2/2