

Name of Laboratory : Laboratory of Environment and Pollution Control Office 10 (Khon Kaen)
Address : Environment and Pollution Control Office 10 (Khon Kaen),
283 Klang Mueang Road, Tambon Nai Mueang, Amphoe Mueang,
Changwat Khon Kaen 40000
Accreditation Number : Testing – 0034
Contact Person : Miss Araya Nuntapotidech
Phone : 0-4324-6772-3
Fax : 0-4324-6772-3, 0-4323-6127
E-mail : envioff@esanenvi.com
Initial Issue Date : 11th June 2009
Issue Date : 14th September 2023
Expired date : 13th September 2027
Standard : ISO/IEC 17025:2017
Reference No. : 0303/15076

Scope of Testing Laboratory Accreditation

Laboratory Name : Laboratory of Environment and Pollution Control Office 10 (Khon Kaen)

Address : Environment and Pollution Control Office 10 (Khon Kaen), 283 Klang Mueang Road,
 Tambon Nai Mueang, Amphoe Mueang, Changwat Khon Kaen 40000

Accreditation Number : Testing - 0034

Laboratory Status : Permanent Site Temporary Mobile

Item Number	Test Material / Product	Test Item / Range of Testing	Test Method / Technique Used
1	Surface water	- Total suspended solids dried from 103 °C to 105 °C 5 mg/L to 500 mg/L	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 24 th ed., 2023, part 2540 D
2	Wastewater	- Total suspended solids dried from 103 °C to 105 °C 10 mg/L to 500 mg/L - COD 40 mg/L to 2 000 mg/L	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 24 th ed., 2023, part 2540 D In - house method : T-7.2-02 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 24 th ed., 2023, part 5220 C

Initial Issue Date 11th June 2009

Issue Number 7

Bureau of Laboratory Accreditation, Department of Science Service, Ministry of Higher Education, Science, Research and Innovation

Scope of Testing Laboratory Accreditation

Laboratory Name : Laboratory of Environment and Pollution Control Office 10 (Khon Kaen)

Address : Environment and Pollution Control Office 10 (Khon Kaen), 283 Klang Mueang Road,
 Tambon Nai Mueang, Amphoe Mueang, Changwat Khon Kaen 40000

Accreditation Number : Testing - 0034

Laboratory Status : Permanent Site Temporary Mobile

Item Number	Test Material / Product	Test Item / Range of Testing	Test Method / Technique Used
3	Tapwater	- Chromium 0.010 mg/L to 0.200 mg/L - Copper 0.100 mg/L to 2.000 mg/L - Iron 0.100 mg/L to 2.000 mg/L - Manganese 0.050 mg/L to 1.000 mg/L - Zinc 0.100 mg/L to 5.000 mg/L	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 24 th ed., 2023, part 3120 B, 3030 K

Initial Issue Date 11th June 2009

Issue Number 7

Bureau of Laboratory Accreditation, Department of Science Service, Ministry of Higher Education, Science, Research and Innovation