

Name of Laboratory : Science Center Laboratory, Faculty of Science and Technology,
Pibulsongkram Rajabhat University

Address : 156 Moo 5, Tambon Plichumpol, Amphoe Mueang,
Changwat Phitsanulok 65000

Accreditation Number : Testing - 0321

Contact Person : Mrs. Wanpen Trongtorkit

Phone : 082-4088216

Fax : 055-267054

E-mail : wanpent.trongtorkit@gmail.com

Website : science.psru.ac.th

Initial Issue Date : 14th June 2024

Issue Date : 14th June 2024

Expired date : 13th June 2028

Standard : ISO/IEC 17025:2017

Reference No. : 0303/8761

Scope of Testing Laboratory Accreditation

Laboratory Name : Science Center Laboratory, Faculty of Science and Technology,
Pibulsongkram Rajabhat University

Address : 156 Moo 5, Tambon Plichumpol, Amphoe Mueang,
Changwat Phitsanulok 65000

Accreditation Number : Testing - 0321

Laboratory Status : Permanent Site Temporary Mobile

Item Number	Test Material / Product	Test Item / Range of Testing	Test Method / Technique Used
1	Drinking water	- pH 6.0 to 9.2	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 24 th ed., 2023, part 4500-H ⁺ B
2	Surface water	- pH 4.0 to 9.2	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 24 th ed., 2023, part 4500-H ⁺ B
3	Effluent water	- pH 4.0 to 9.2	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 24 th ed., 2023, part 4500-H ⁺ B

Initial Issue Date 14th June 2024

Issue Number 1

Bureau of Laboratory Accreditation, Department of Science Service, Ministry of Higher Education, Science, Research and Innovation